



California Medical Device Recall Information



Recall Name

Stryker Recalls Fuhrman Pleural and Pneumopericardial Drainage Sets Due to the Potential that the Catheter May Break During Insertion

Recall Date	Product Description	Recalling Firm	Recall Reason
01/11/16	Fuhrman Pleural and Pneumopericardial Drainage Sets Item Number: G03974	Stryker Corporation Tempe, AZ	<i>Potential for the catheter included in the Drainage Set to break off in the pleural cavity.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Suspect Lot Numbers recalled: <ul style="list-style-type: none">• 295397U• 331621U• 374044U• 394431U• 406125U• 406131U• 410982U• 412655U• 415094U• 422060U• 519812U• 7718289U• 838193U	CA , nationwide	Manufactured between: November 6, 2009 and October 21, 2011

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm481231.htm>